

Patient: AHMED KAROICHE (male, *1958-01-01, #RF67330)
Series: Dose SR (#10002)
Manufacturer: UIH (uCT 530, #300183)
Preliminary Flag: FINAL
Completion Flag: COMPLETE
Verification Flag: UNVERIFIED
Content Date/Time: 2025-08-14 11:36:49

X-Ray Radiation Dose Report

Concept Modifier: Procedure reported [[Annex 1](#)]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 300183
Observation Context: Start of X-Ray Irradiation = 2025-08-14 11:13:33 +00:00
Observation Context: End of X-Ray Irradiation = 2025-08-14 11:36:39 +00:00
Observation Context: Scope of Accumulation [[Annex 2](#)]

Source of Dose Information:
Automated Data Collection (113856, DCM)

CT Accumulated Dose Data

Total Number of Irradiation Events:
5 {events}
CT Dose Length Product Total:
1728.87 mGy.cm

CT Acquisition

Acquisition Protocol:
Scout
Target Region:
Abdomen (T-D4000, SRT)
CT Acquisition Type:
Constant Angle Acquisition (113805, DCM)
Irradiation Event UID:
1.2.156.112605.189250942751537.250814105501.9.7020.32585

CT Acquisition Parameters

Exposure Time:
3.538665 s
Scanning Length:
640.0111 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
2.2 mm

Number of X-Ray Sources:

1 {X-Ray sources}

CT X-Ray Source Parameters**Identification of the XRay Source:**

A

KVP:

120 kV

Maximum X-Ray Tube Current:

67.41573 mA

X-Ray Tube Current:

40 mA

CT Dose**Mean CTDIvol:**

0.03627 mGy

CTDIw Phantom Type:

IEC Body Dosimetry Phantom (113691, DCM)

DLP:

2.33 mGy.cm

CT Acquisition**Acquisition Protocol:**

Arterial Phase

Target Region:

Abdomen (T-D4000, SRT)

CT Acquisition Type:

Spiral Acquisition (P5-08001, SRT)

Irradiation Event UID:

1.2.156.112605.189250942751537.250814111418.9.7020.56021

CT Acquisition Parameters**Exposure Time:**

18.35378 s

Scanning Length:

596.4703 mm

Nominal Single Collimation Width:

0.55 mm

Nominal Total Collimation Width:

22 mm

Pitch Factor:

1.075 {ratio}

Number of X-Ray Sources:

1 {X-Ray sources}

CT X-Ray Source Parameters**Identification of the XRay Source:**

A
KVP:
120 kV
Maximum X-Ray Tube Current:
379.6613 mA
X-Ray Tube Current:
179.6786 mA
Exposure Time per Rotation:
0.7 s

CT Dose

Mean CTDI_{vol}:
9.798656 mGy
CTDI_w Phantom Type:
IEC Body Dosimetry Phantom (113691, DCM)
DLP:
608.23 mGy.cm

CT Acquisition

Acquisition Protocol:
Portal Phase
Target Region:
Abdomen (T-D4000, SRT)
CT Acquisition Type:
Spiral Acquisition (P5-08001, SRT)
Irradiation Event UID:
1.2.156.112605.189250942751537.250814111418.9.7020.103763

CT Acquisition Parameters

Exposure Time:
12.90639 s
Scanning Length:
412.4145 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
22 mm
Pitch Factor:
1.075 {ratio}
Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:
A
KVP:
120 kV
Maximum X-Ray Tube Current:

379.8648 mA

X-Ray Tube Current:

185.8214 mA

Exposure Time per Rotation:

0.7 s

CT Dose

Mean CTDI_{vol}:

10.13219 mGy

CTDI_w Phantom Type:

IEC Body Dosimetry Phantom (113691, DCM)

DLP:

442.29 mGy.cm

CT Acquisition

Acquisition Protocol:

TARDIF

Target Region:

Abdomen (T-D4000, SRT)

CT Acquisition Type:

Spiral Acquisition (P5-08001, SRT)

Irradiation Event UID:

1.2.156.112605.189250942751537.250814111630.9.7020.118408

CT Acquisition Parameters

Exposure Time:

12.90849 s

Scanning Length:

412.421 mm

Nominal Single Collimation Width:

0.55 mm

Nominal Total Collimation Width:

22 mm

Pitch Factor:

1.075 {ratio}

Number of X-Ray Sources:

1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:

A

KVP:

120 kV

Maximum X-Ray Tube Current:

380.0073 mA

X-Ray Tube Current:

185.8214 mA

Exposure Time per Rotation:

0.7 s

CT Dose

Mean CTDI_{vol}:

10.15305 mGy

CTDI_w Phantom Type:

IEC Body Dosimetry Phantom (113691, DCM)

DLP:

443.43 mGy.cm

CT Acquisition

Acquisition Protocol:

TARDIF D.V

Target Region:

Abdomen (T-D4000, SRT)

CT Acquisition Type:

Spiral Acquisition (P5-08001, SRT)

Irradiation Event UID:

1.2.156.112605.189250942751537.250814113450.9.7020.21321

CT Acquisition Parameters

Exposure Time:

7.518045 s

Scanning Length:

230.424 mm

Nominal Single Collimation Width:

0.55 mm

Nominal Total Collimation Width:

22 mm

Pitch Factor:

1.075 {ratio}

Number of X-Ray Sources:

1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:

A

KVP:

120 kV

Maximum X-Ray Tube Current:

282.9344 mA

X-Ray Tube Current:

167.3929 mA

Exposure Time per Rotation:

0.7 s

CT Dose

Mean CTDIvol:

9.134122 mGy

CTDIw Phantom Type:

IEC Body Dosimetry Phantom (113691, DCM)

DLP:

232.59 mGy.cm

Annex

[Annex 1](#)

Procedure reported:

Computed Tomography XRay (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

[Annex 2](#)

Scope of Accumulation:

Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.4892943.343.20250814104940.2692.73832

This page was generated from a DICOM Structured Reporting document by [OFFIS DCMTK](#) 3.6.5.